

Precision digital pressure gauge - Model CPG1500

Precision digital pressure gauge - Model CPG1500

Item number: 52785803

Quantity: 1 pcs.

Order code: CPG1500-ST-Z-S-BV325-NBSZ-53-CZHZ-Z-Z2ZZ W
 CERT-PR-DE1-AP03C1B-Z-C12-ZZ-Z 1
 CERT22-EGCP-BB- CD

Overview: Precision digital pressure gauge - Model CPG1500
 Specifications according to data sheet: CT 10.51
 explosion protection: without
 Pressure range: -1...+1,5 bar gauge
 Process Connection: 1/4 NPT
 wetted parts: stainless steel
 Accuracy: 0.05 % FS
 Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204 (factory calibration)
 Communication: Bluetooth®
 Info Bluetooth®: Please bear in mind the relevant radio approvals in your country.
 incl. protective housing cap

Calibration of pressure measuring instruments

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204 (factory calibration)
 Type of device under test: Electrical measuring device with display (digital pressure gauge, pressure controller, etc.)
 Accuracy: > 0,02 % ... < 0,1 %
 Numbers of measurement series: 2
 Number of measuring points per measurement series: 9
 Pressure type - pressure: Relative pressure (over- or underpressure), Gauge or Vacuum Pressure
 Calibration range - pressure: > 10 mbar ... <= 400 bar

Documentation: Calibration (documentation of the values)
 1. Language of certificate: DE
 2. Language of certificate: GB

Test report 2.2 per EN 10204
 Certificate Language: English and German
 Certificate for: Digital pressure gauge
 Material certificate: Wetted metallic components per NACE MR 0175

Date: 8/28/2023



Details:

Precision digital pressure gauge - Model CPG1500

Specifications according to data sheet: CT 10.51

explosion protection: without

Pressure range: -1...+1,5 bar gauge

Process Connection: 1/4 NPT

wetted parts: stainless steel

Accuracy: 0.05 % FS

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204 (factory calibration)

Communication: Bluetooth®

Info Bluetooth®: Please bear in mind the relevant radio approvals in your country. incl. protective housing cap

Calibration of pressure measuring instruments

Type of certificate: Acceptance test certificate 3.1 in accordance with DIN EN 10204 (factory calibration)

Type of device under test: Electrical measuring device with display (digital pressure gauge, pressure controller, etc.)

Accuracy: > 0,02 % ... < 0,1 %

Numbers of measurement series: 2

Number of measuring points per measurement series: 9

Pressure type - pressure: Relative pressure (over- or underpressure), Gauge or Vacuum Pressure

Calibration range - pressure: > 10 mbar ... <= 400 bar

Documentation: Calibration (documentation of the values)

1. Language of certificate: DE

2. Language of certificate: GB

Test report 2.2 per EN 10204

Certificate Language: English and German

Certificate for: Digital pressure gauge

Material certificate: Wetted metallic components per NACE MR 0175